



US006146331A

United States Patent [19]

Wong

[11] **Patent Number:** **6,146,331**[45] **Date of Patent:** **Nov. 14, 2000**

[54] **METHOD FOR IMPROVED CLUTTER SUPPRESSION FOR ULTRASONIC COLOR DOPPLER IMAGING**

[75] Inventor: **Tommy Wong**, Issaquah, Wash.

[73] Assignee: **Siemens Medical Systems, Inc.**, Iselin, N.J.

[21] Appl. No.: **09/164,386**

[22] Filed: **Sep. 30, 1998**

[51] **Int. Cl.⁷** **A61B 8/00**

[52] **U.S. Cl.** **600/454**

[58] **Field of Search** 600/454-456

[56] **References Cited**

U.S. PATENT DOCUMENTS

4,683,893	8/1987	Mayo	128/660
4,850,364	7/1989	Leavitt	128/661.09
4,896,674	1/1990	Seo	128/661.09
5,243,987	9/1993	Shiba	128/660.06
5,341,808	8/1994	Rickey et al.	128/660.01
5,349,525	9/1994	Dunki-Jacobs et al.	364/413.25
5,494,037	2/1996	Banjanin et al.	600/455
5,522,393	6/1996	Phillips et al.	600/455
5,553,621	9/1996	Otterson	600/455
5,653,234	8/1997	Kim et al.	128/660.01
5,664,571	9/1997	Yamazaki	600/455
5,706,817	1/1998	Song et al.	128/661.09

5,706,819	1/1998	Hwang et al.	128/662.02
5,709,209	1/1998	Friemel et al.	128/660.07
5,709,210	1/1998	Green et al.	128/661.07
5,740,806	4/1998	Miller	128/661.01
5,913,824	6/1999	Ogasawara et al.	600/455

OTHER PUBLICATIONS


"Bias and Variance in the Estimate of the Doppler Frequency Induced by a Wall Motion Filter," J. C. Willemetz, et al., taken from Ultrasonic Imaging, vol. 11, pp. 215-225 (1989).
"On the Performance of Regression and Step-Initialized IIR Clutter Filters for Color Doppler Systems in Diagnostic Medical Ultrasound," by Anthony P. Kadi & Thanasis Loupas, IEEE Transactions on Ultrasonics, Ferroelectrics and Frequency Control, vol. 42, No. 5, Sep. 1995, pp. 927-937.



Primary Examiner—Francis J. Jaworski

[57] **ABSTRACT**

A Doppler ultrasound clutter suppression system and method in which the clutter is first low-pass filtered (716) to suppress the flow component (4024) just before mean frequency estimation (718). The mean frequency is then estimated and mixed with the original clutter data which is positioned at DC. The low-pass filter bandwidth is predetermined based on the particular application and imaging parameters. The signal is then high-pass filtered (726) to remove the clutter component (4022).

14 Claims, 6 Drawing Sheets








FIG. 3




FIG. 4




FIG. 5




FIG. 6(a)




FIG. 6(b)




FIG. 6(c)




FIG. 7

METHOD FOR IMPROVED CLUTTER SUPPRESSION FOR ULTRASONIC COLOR DOPPLER IMAGING

BACKGROUND OF THE INVENTION

1. Field of the Invention

The present invention relates to an ultrasonic imaging system, and more particularly, to a filtering method for removing a low-frequency Doppler signal from an ultrasound image.

2. Description of the Related Art

Medical imaging systems are used as diagnostic tools for viewing internal areas of the patient's body. Imaging systems based on ultrasound technology are desirable because such systems are non-invasive, portable and generally easy to use. A Doppler ultrasound system displays a frequency deviation amount together with a scattering strength of the signal as image on a screen to enable evaluation of a dynamic function of a living body. In particular, the Doppler imaging system is capable of displaying color images and demodulates a received signal and digitizes the demodulated signal, thereby imaging a flow of blood flowing through a heart or blood vessel into a real-time two-dimensional image. Such color Doppler imaging systems can display cross-sectional images and blood flow information simultaneously. To discriminate the layer image and the blood flow information, the cross-sectional image is indicated as white and black and the blood flow information is indicated as color.

Although an ultrasound beam is focused, some part of the Doppler sample volume typically lies outside the blood vessel. This results in the ultrasound instrument receiving a signal from the vessel wall and surrounding tissue (referred to as the clutter component), as well as from the blood flowing in the vessel (referred to as the flow component). The clutter component returned from the wall and tissue has an amplitude that is generally many times greater than the flow component. Because of respiration and cardiac motion, the tissue and vessel tend to move at slower velocities than the flowing blood. The Doppler signal also contains noise produced by the Doppler instrument's electronics. As such, the Doppler signal is a composite signal composed of a clutter component, a flow component and a noise component. The flow component tends to have a low amplitude and high-frequency, while the clutter component tends to have a large amplitude and low-frequency. In order to ensure clinical efficiency, Doppler instruments must be able to determine the velocity of moving blood accurately in the presence of clutter and noise signal components. Accordingly, the clutter component is generally removed from the composite Doppler signal via filtering prior to velocity estimation. If the clutter component is not filtered, the velocity measurement returns to the velocity of the moving tissue and not of the moving blood. Typically, a high-pass filter referred to as a wall filter is used to remove the clutter component of the signal while leaving the flow component intact.

For example, conventional clutter processing uses a mean frequency method to estimate the clutter frequency. The clutter frequency is then mixed with the original clutter data in an attempt to position the clutter at a predetermined DC level. The conventional method assumes that flow has little effect on the mean frequency estimation. However, this assumption is not valid when the flow component is comparable to the clutter component in amplitude which may occur, for example, in the middle of a big vessel where the effect of clutter is much smaller than at the vessel boundary

and/or where contrast agents are used to boost flow amplitude. This may cause the mean frequency clutter frequency to shift towards the blood flow frequency. When flow velocity is high, the mean frequency shifts toward the flow frequency even more. When this frequency is mixed with the clutter, a combination of the following results: (1) clutter is not positioned at the predetermined DC level resulting in incomplete clutter cancellation; (2) clutter may be even further from the predetermined DC level than it would have been without mixing; and (3) strong blood flow can be misinterpreted as clutter and thus shifted erroneously close to the predetermined DC level, resulting in either incorrect velocity estimation or cancellation by a subsequent wall filter. Conventionally, a combination of clutter mean frequency, power and variance may be used in an attempt to identify such situations and disable mixing accordingly. Although this may mitigate the effect of the strong blood flow being misinterpreted as clutter, clutter residue may still be excessive due to the inaccurate positioning of the clutter away from the predetermined DC level.

Accordingly, there is a need for an improved method for eliminating clutter components from Doppler ultrasound signals. There is a further need for a method for removing clutter components from Doppler ultrasound signals when the flow component is comparable in amplitude to the clutter component. There is further a need to remove the clutter component from the flow component when flow velocity is relatively high.

SUMMARY OF THE INVENTION

These problems in the prior art are overcome in large part by a Doppler ultrasound clutter suppression system and method according to the present invention. In particular, in the present invention the clutter is first low-pass filtered to suppress the flow component just before mean frequency estimation. The mean frequency is then estimated and mixed with the original clutter data and positioned at a predetermined DC level. The low-pass filter bandwidth is predetermined based on the particular application and imaging parameters. When the imaging parameters are altered, the filter bandwidth similarly is altered to compensate for the change in clutter bandwidth.

According to one embodiment, the low-pass filter is a regression-type digital filter. It is noted, however, that other types of filters, such as IIR or FIR filters are contemplated. Moreover, the low-pass filter may be embodied as an analog filter.

BRIEF DESCRIPTION OF THE DRAWINGS

A better understanding of the present invention is obtained when the following detailed description is considered in conjunction with the following drawings in which:

FIG. 1 is a diagram illustrating an exemplary clutter suppression system according to the prior art;

FIG. 2 is a diagram illustrating an exemplary clutter suppression system according to an embodiment of the present invention;

FIG. 3A and FIG. 3B are an exemplary color spectrum before and after being low-pass filtered;

FIG. 4A and FIG. 4B illustrate the input and output of a wall filter in the case in which there has been no shifting;

FIG. 4C and FIG. 4D illustrate the input and output of a wall filter in the case in which there has been spectrum shifting and clutter mean frequency estimation;

FIG. 4E and FIG. 4F illustrate the input and output of a wall filter in the case in which there has been spectrum shifting and low-pass filtering with clutter mean frequency suppression;

FIG. 5 illustrates an exemplary ultrasound system employing clutter suppression;

FIGS. 6A, 6B and 6C illustrates exemplary dynamic bandwidth control; and

FIG. 7 is a flowchart of operation of an ultrasound system according to an embodiment of the invention.

DETAILED DESCRIPTION OF THE INVENTION

Turning now to the drawings, and with particular attention to FIG. 5, a block diagram of an exemplary ultrasound imaging system is shown. An ultrasound transmitter **410** transmits ultrasound energy through an ultrasound transducer **412**, which typically includes an array of ultrasound transducer elements. The ultrasound energy is transmitted into a target region of a patient. A portion of the transmitted ultrasound energy is reflected by various components of the target region. Reflected ultrasound energy is received by the transducer **412** and is input to a receiver **414**, which performs an analog-to-digital conversion and focuses the signals from the multiple transducer elements.

The signals from the receiver **414** are input to a black and white processor **416** and to a base band converter **418**. The base band converter **418** converts the RF frequencies to base band and inputs the base band signals to the black and white processor **416** and to a mean frequency estimator **422**. The mean frequency estimator **422** may include further frequency shifting to position the clutter signal components at DC and low-pass filtering according to the present invention. A control processor **421** coupled to a memory **423** may be provided to receive user inputs and adjust bandwidths of the low-pass filter and the wall filter, as will be discussed in greater detail below. The samples output from the mean frequency estimator **422** are provided to a wall filter **424**. As noted previously, in color Doppler imaging systems, stationary and slowly moving targets produce large, low frequency signals. These unwanted signals, positioned accurately at DC by the mean frequency estimator **422** according to the present invention, are removed by the wall filter **424** to detect the higher frequency Doppler blood echoes. Outputs of the wall filter **424** are input to a velocity estimation unit **426** which makes an estimate of the velocity of blood in the target region.

The velocity estimate and the output of the black and white processor **416** are input to a scan converter **428** which converts the color and black and white data from acoustic space to raster space coordinates. A color map unit **430** maps black and white data and velocities output by the scan converter **428** to color and brightness, which are displayed on a TV monitor **432**. An exemplary ultrasound system including a wall filter **424**, a black and white processor **416**, a scan converter **428**, and a color map unit **430** is the Elegra, available from Siemens Medical Systems, Inc.

It is noted that, while the analog to digital conversion is illustrated as occurring before base band conversion and the mean frequency estimation, the analog to digital conversion may equally be performed after it, and in fact, may be performed after wall filtering. Thus, FIG. 5 is exemplary only.

A block diagram showing an exemplary clutter suppression mechanism or mean frequency estimation block **420** according to the prior art is shown in FIG. 1. In particular, the mean frequency estimator **100**, may correspond generally to the mean frequency estimator **420** of FIG. 5, and includes a frequency estimator **102**, which receives the received base band Doppler signal. The received base band

Doppler signal is also provided to a mixer **104**. The output of the frequency estimator **102** is mixed with the received base band Doppler signal input. The output of the mixer **104** is assumed to position the clutter signal at DC. As discussed above, the mean frequency estimator **100** may further include one or more digital signal processors (DSP) **101** for computing clutter mean velocity, power and variance to disable mixing under certain circumstances, discussed above. This is represented schematically by a switch **103**. It is noted that the mean frequency estimation and the DSP functions described above are typically accomplished using the same DSP; thus, the FIG. 1 is exemplary only. If is further noted that the DSP operations may occur in parallel with the mean frequency estimation. The mixed signal (or the bypass signal) is then provided to a wall filter such as the wall filter **424** of FIG. 5. The wall filter **424** is typically a high-pass filter to eliminate the low frequency clutter components. However, as discussed above, this system is disadvantageous in that it cannot account for the case in which the flow signal is of similar magnitude to the clutter signal.

Accordingly, with reference to FIG. 2, an improved clutter suppression system or mean frequency estimation block **200**, according to the present invention, is illustrated. Again, the mean frequency estimation block **200** corresponds generally to the mean frequency estimator **420** of FIG. 5. In this case, however, the received base band Doppler signal is input to a low-pass filter **202**. The low-pass filter **202** may be any low-pass filter adequate to filter the flow component of the color Doppler signal. In particular, the low-pass filter **202** may be a digital regression-type filter of the type described in Kadi, et al., "On the Performance of Regression and Step-Initialized IIR Clutter Filters for Color Doppler Systems in Diagnostic Medical Ultrasound," IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control, Vol. 42, No. 5, Sep. 1995, which is hereby incorporated by reference in its entirety as if fully set forth herein. It is noted, however, that other types of low-pass filters may be employed.

The output of the low-pass filter **202** is provided to a DSP or other processing or control unit **205**, as will be described below. The DSP **205** may cause the signal to be provided to a frequency estimator **204**, and then may also be provided to a mixer **206**. The output of the frequency estimator **202** is mixed with the received Doppler signal input. The mixed signal is then provided to a wall filter such as the wall filter **424** of FIG. 5. The wall filter **106** is typically a high-pass filter to eliminate the low frequency clutter components. Alternatively, the DSP **205** may cause the received base band signal to be provided directly to the wall filter.

As in the example shown in FIG. 1, the DSP **205** may be provided to receive the low pass filter output for mean velocity, power and variance calculations. If one or more predetermined conditions are met, no mixing and mean frequency of estimation may occur. For example, if the power level is less than a predetermined threshold, no shifting will occur. This may be the case, for example, where there is a flow component, but no clutter component. In that case, the clutter power is compared to the noise level. If the clutter power is below noise, then no mixing is permitted.

An exemplary response characteristic for the low-pass filter **202** is shown in FIG. 3A and 3B. In particular, FIG. 3A illustrates a color spectrum **1010** sampled close to the center of a carotid artery. As can be seen, the color spectrum includes a clutter peak **1000** and a flow peak **1002**. In the example shown, the flow peak **1002** is actually higher in magnitude than the clutter peak **1000**.

FIG. 3B illustrates the color spectrum **1010** after low-pass filtering has occurred. As can be seen, the clutter peak **1000**

is still prominent, but the flow peak **1002** has been almost completely eliminated. The low-pass filter **102** thus compensates for the prominent flow peak **1002**. The output of the low-pass filter **102** is then provided to the mean frequency estimator **204**, as discussed above.

FIG. **4** illustrates the inputs and outputs of the wall filter **424** under various test circumstances. In particular, FIG. **4A** and **4B** illustrate the input and output of wall filter **424** in the case in which there has been no clutter shifting, i.e., there has been no estimate of the clutter mean frequency and attempt to position the clutter at DC. As can be seen in FIG. **4A**, the color spectrum **4000a** displays a prominent clutter peak **4008a** as well as a prominent flow peak **4006a**. The clutter peak **4008a** is positioned away from DC. FIG. **4B** illustrates the spectrum **4000b** after it has been high-pass filtered, i.e., at the output of the wall filter **424**. The spectrum **4000b** thus shows a prominent color flow peak **4006b**. The clutter peak **4008a** has been generally eliminated, though a residue peak **4000b** remains.

FIGS. **4C** and **4D** show the input and output, respectively, of the wall filter **424** after the clutter spectrum has been shifted and with clutter mean frequency estimation performed thereon. As can be seen in FIG. **4C**, the color spectrum **4010a** includes a prominent clutter peak **4016a** and a flow peak **4014a**. In this case, the clutter peak **4016a** is actually farther away from DC than in the non-clutter-shifted case.

The output of the wall filter **424** for this case is shown in FIG. **4D**. In particular, the wall-filtered spectrum **4010b** includes both a prominent clutter peak **4016b** and a prominent flow peak **4014b**. In this case, however, the magnitude of the clutter peak **4016b** is significantly higher than that of the flow peak **4014b**. Consequently, any velocity estimates would necessarily show a sub-optimal result. However, as will be seen with regard to FIG. **4E** and FIG. **4F**, this problem is overcome by the present invention.

More particularly, FIG. **4E** and FIG. **4F** illustrate the input and output spectra **4020a**, **4020b**, respectively, of the wall filter **424** in the case in which the color spectrum is low-pass filtered with clutter mean frequency estimation and shifting. As can be seen in FIG. **4E**, the clutter peak **4022a** of the color spectrum **4020a** is positioned almost exactly at DC. The color flow peak **4024a** is positioned substantially away from DC. The color spectrum is then high-pass filtered using the wall filter **424** as seen in FIG. **4F**. The resulting color spectrum **4020b** has a prominent color flow peak **4024b** and no clutter peak. Thus, as can be appreciated, the use of the low-pass filter in accordance with the present invention can provide for more accurate clutter removal than standard mean frequency estimation and shifting.

As discussed above, the control processor **421** may be used to dynamically control the low-pass filter bandwidth, depending on test or scan circumstances. For example, a scan of the aorta might require a different filter bandwidth than a scan of the carotid artery. Similarly, a scan at one pulse repetition frequency (PRF) might require a different low-pass filter bandwidth. (The bandwidth of the wall filter may similarly be adjusted). More particularly, the control processor **421** is configured to receive inputs from a sonographer (typically via a keyboard (not shown)), which control scan parameters, such as PRF and the like. A memory **423** coupled to the control processor **421** may include a look-up table defining filter bandwidths for the particular scan. Depending on the input scan parameters, the control processor **421** may access a low-pass filter from the memory **423** for use by the mean frequency estimation unit **422**.

For example, FIG. **6A** illustrates a color spectrum **600a** having a flow component **604a** and a clutter component **602a**. A low-pass filter **606a**, having exemplary parameters as illustrated by the dashed line, may be used to remove the flow component **604a** prior to mean frequency estimation. In FIG. **6B**, a color spectrum **600b** is shown which represents the case in which the sonographer increases the pulse repetition frequency. The color spectrum **600b** has a flow component **604b** and a clutter component **602b**. If the same filter **606a** as used in FIG. **6A** is employed, the filter is ineffective to remove the flow component **604b**. However, if the bandwidth of the filter is changed, as shown in FIG. **6C**, then the flow component may be adequately suppressed.

Finally, it is noted that, while the examples discussed involve shifting to DC, it is possible to perform the clutter filtering without shifting the clutter to DC. However, this is less desirable since a more complex clutter filter is required.

Turning now to FIG. **7**, a flowchart **700** illustrating general operation of an ultrasound system according to an embodiment of the present invention is shown. In particular, in a step **702**, the ultrasound system receives scan parameters. The scan parameters are typically received, for example, via keyboard input and processed by the control processor **421**. Scan parameters can include, for example, pulse repetition frequencies and other known parameters such as volume to be scanned including the specific target area such as the aorta, carotid artery and the like. In a step **704**, the control processor **421** accesses one or more lookup tables in the memory **423**. As noted above, the lookup tables may include bandwidth and filter parameters for use by the mean frequency estimation unit **422** and/or the clutter filter **424**. In a step **706**, the control processor **421** adjusts the filters by instructing the mean frequency estimation unit **422** and/or the clutter filter **424** with the new filter parameters. In a step **708**, the sonographer performs the scan on the patient over the designated region. As is well known, the transmitter **410** causes pulses to be emitted from the transducer **412**. Received pulses are received at the transducer and the receiver **414** in a step **710**. In a step **712**, the receiver **414** converts the received signals into digital format. In a step **714**, the base band converter **418** converts the received signals down to base band. Next, in a step **716**, the low pass filter receives the baseband signal and suppresses the flow component for mean frequency estimation. In a step **718**, the mean frequency estimation unit **422** performs a power spectrum calculation on the received, low pass filtered signal. If, in a step **720** the power spectrum magnitude is determined to be less than a predetermined threshold, then in a step **726**, the baseband signal bypasses the mean frequency estimation and mixing, and, in a step **726**, a wall filtering is performed using the clutter filter **424**. Then the velocity estimation unit **726** estimates the velocity in a step **728** and the resulting output is displayed in a step **730**.

If, however, in step **722** the power spectrum was greater than a predetermined threshold (such as when clutter is present in significant amounts) then the result of the mean frequency estimation is provided to the mixing unit and, in a step **724**, the estimated frequency is mixed with the ultrasound signal to position the clutter frequency at DC. The resulting mixed signal is then provided to the wall filter **424**, in a step **726**, and the velocity estimation unit **426**, in a step **728**. The ultrasound signal is then provided to a display in a step **730**.

What is claimed is:

1. An ultrasound imaging system, comprising:

means for receiving an ultrasound signal, said ultrasound signal including a clutter component and a flow component;

7

means coupled to said receiving means for suppressing said flow component, said suppressing means including a low-pass filter for suppressing said flow component; and

means for frequency shifting and mean frequency estimating a resulting low-pass filtered clutter component.

2. The ultrasound imaging system of claim 1, said suppressing means including a means for high-pass filtering said ultrasound signal after said mean frequency estimation and frequency shift.

3. The ultrasound imaging system of claim 1, wherein said low-pass filter is a regression-type filter.

4. The ultrasound imaging system of claim 1, wherein said low-pass filter is an analog filter.

5. A method for processing ultrasound signals, comprising:

receiving an ultrasound signal including a flow component and a clutter component;

filtering said ultrasound signal to remove said flow component during a frequency estimation;

using an output of said frequency estimation to position said ultrasound signal such that said clutter component is substantially at a predetermined frequency; and

filtering said ultrasound signal to remove said clutter component at said predetermined frequency.

6. A method according to claim 5, wherein said filtering to remove said flow component is digital low-pass filtering.

7. A method according to claim 5, wherein said filtering to remove said flow component is analog low-pass filtering.

8. A method according to claim 6, wherein said digital low-pass filtering is regression filtering.

8

9. An ultrasound imaging system, comprising:

one or more transducers for transmitting ultrasound energy into a target region of a patient;

a receiver configured to receive reflected ultrasound energy and providing a received ultrasound signal, said received ultrasound signal including a clutter component and a flow component; and

a clutter suppression system, said clutter suppression system including:

a first filter for suppressing said flow component from said ultrasound signal;

a frequency estimator, responsive to said ultrasound signal having said flow component suppressed, for determining a frequency estimate of said clutter component; and

a second filter for removing said clutter component from said ultrasound signal after said frequency estimate of said clutter component has been made.

10. An ultrasound system according to claim 9, wherein said first filter is a low-pass regression filter.

11. An ultrasound system according to claim 9, wherein said second filter is a high-pass filter.

12. An ultrasound system according to claim 1, further including means for dynamically changing a bandwidth of said low-pass filter depending on scan parameters.

13. An ultrasound system according to claim 12, further including means for dynamically changing a bandwidth of said high-pass filter depending on said scan parameters.

14. An ultrasound system according to claim 1, wherein said suppressing means includes means for bypassing said frequency shift of a power spectrum of said clutter component is less than a predetermined threshold.

* * * * *